Shutdown works: new BSToF detector

Will replace the radiation monitor (S.Ghazaryan and A.Hovhannisyan)

One scintillator plane, d=2cm,8-fold ϕ segmentation, Spacal-type Hamamatsu PMs. The readout and HV will be integrated into the TOF readout system:

- 'OR' of all channels will be fed to the same place as old rad.monitor
- For each channel the rate and timing histograms will be shown in the TOF monitor





BSToF is installed and tested

